

检 测 报 告

报告编号: LH2023030005

第 1 页 共 4 页

产 品 名 称: 智能吸氢机

委 托 单 位: 中韩大字(深圳)实业发展有限公司

生 产 单 位: 中韩大字(深圳)实业发展有限公司

检 测 类 别: 委托检测

报 告 日 期: 2023.03.10

检 测 机 构 名 称: 广东量 化 检 测 技 术 有 限 公 司

(检验检测专用章)

地址: 广东省佛山市南海区狮山镇广云路仙溪段广东生物医药产业基

地一期第一组团C栋五层自编501号

电话: 0757-29956220

传真: 0757-29956220

邮编: 528225

电子邮件: lhjcgd@163.com

检测报告

报告编号: LH2023030005

第 2 页 共 4 页

样品名称	智能吸氢机	产品型号	/
		样品数量	1 台
委托单位	中韩大字（深圳）实业发展有限公司	检测类别	委托检测
		收样日期	2023-03-07
来样方式	送样检测	检测日期	2023-03-10
送样人	/	检讫日期	2023-03-10
检测项目	氢气纯度、氧气纯度、氢氧比例		
检测依据	参照 CB 31633-2014《食品添加剂 氢气》 GB 8982-2009《医用及航空呼吸用氧》		
主要检测设备	气相色谱仪（QTT-ZJ-070） 氧分析仪（QTT-ZJ-700）		
检测结论	见检测结果内页 广东量化检测技术有限公司 (检验检测专用章) 批准日期: 2023 年 03 月 10 日		
检测环境条件	温度: 22.5℃	相对湿度: 50%RH	大气压: 101.1kPa
备注			

主 检: 邱惠如

邱惠如

审 核: 钟庚

钟庚

批 准: 温晓辉

温晓辉

检 测 报 告

报告编号: LH2023030005

第 3 页 共 4 页

序号	检测项目	单位 符号	检测方法	检测结果
1	氢气纯度	% (mol/mol)	GB/T31633-2014《食品添 加剂 氢气》 GB 8982-2009《医用及航 空呼吸用氧》	99.99
2	氧气纯度	% (mol/mol)		99.99
3	氢气: 氧气	% (mol/mol)		66.7: 33.3

检测样品照片如下:



结束

技术审核
用章

注 意 事 项:

Matters Needing Attention

1. 报告无“检验检测专用章”、检验单位骑缝章无效。
2. 复制报告未重新加盖“检验检测专用章”、检测单位骑缝章无效。
3. 报告无主检、审核、批准人签字无效。
4. 报告涂改无效。
5. 委托单位或生产单位对检测报告如有异议,应在收到检测报告后的 15 日内向检验机构提出,逾期不予受理。
6. 委托人送检样品,检验结果仅对来样负责。
7. 本报告仅供内部了解品质使用,不做其它公正性等法律用途。
1. It is invalid without the “Special seal for testing” or the inspection official seal.
2. The copy of the report is invalid without new affxing seal of the “Special seal for testing” or the inspection official seal.
3. The report is invalid without the signatures of major inspector, verifier and approver.
4. The report is invalid if it is altered.
5. Objections to the inspection report submitted to the inspection organization within 15(rirteen)days from the date of receipt of the report. Being overdue is inadmissible.
6. The entrusted inspection organization is only responsible for the sent samples.
7. This report is only for internal understanding of quality, not for other legal purposes such as fairness.

检测机构简介:

广东量化检测技术有限公司为独立的第三方检测校准机构,并获得了 CNAS 实验室认可和 CMA 检验检测机构资质认定。业务范围覆盖工业气体检测、洁净室检测、电子特气管道五项检测、仪器设备计量校准。

实验室配备了国际一流品牌的检测设备,检测人员皆来自国内外知名院校和科研机构,关键人员从业经历二十年以上。同国内科研机构和大专院校的实验室建立了技术协作共享,检测结果在行业内获得广泛认可。